



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204096	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 02/28/2014			
LOCATION OF INSTRUMENT (STREET AND CITY) Marion Co. Sheriff's Dept., 1703 Marion City Rd., Palmyra, MO 63461		TIME OF INSPECTION 2:06 pm			
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.					
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 02/28/2014/ 14:06			
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR				
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS				
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD				
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION				
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER				
<input checked="" type="checkbox"/> INDICATOR LIGHTS					
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 12002 EXP. DATE 08/29/2014					
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN 1709 EXP. DATE 04/29/2014					
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE					
TEST 1 ➡ .096	TEST 2 ➡ .098	TEST 3 ➡ .098			
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)					
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)					
REFUSALS 0	(0-.04) 0	(.05-.09) 2	(.10-.14) 1	(.15-.19) 0	OVER .19 0
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY). Instrument meets Dept. of Health specs.					
INSPECTING OFFICER					
SIGNATURE 			PRINT FULL NAME Brandon C. Griffin		
TYPE II PERMIT NUMBER/EXPIRATION DATE 240019 01/28/2016			TELEPHONE NUMBER (660) 385-2132		
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901					

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRANDON C GRIFFIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/28/2014

NUMBER 240019

EXPIRES 1/28/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MD 580-0771 (6-10)

LAB-4 (R6-10)

	STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM
	INSTRUMENT OPERATOR CARD
<small>The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.</small>	
Operator GRIFFIN, BRANDON Permit No 240019 Date Issued 1/28/2014 Date Expires 1/28/2016	

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204096
02/28/14

TESTING OFFICER:

GRIFFIN/B/C

OFFICER I.D.: 0429

PERMIT NUMBER: 240019

EXPIRATION DATE: 01/28/16

MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	14:12
INTERNAL STANDARD	VERIFIED	14:12
EXTERNAL STANDARD	.096	14:13
BLANK TEST	.000	14:13
EXTERNAL STANDARD	.098	14:14
BLANK TEST	.000	14:14
EXTERNAL STANDARD	.098	14:15
BLANK TEST	.000	14:15

N = 3

SUN. = .1

MS. = .0979

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204096
02/28/14

ARREST TIME: 12:00

SUBJECT NAME:

TEST

DOB: 01/01/76

SEX: M

STATE/O.C.L.: MO/1234567890

ARRESTING OFFICER:

GRIFFIN/B/C

OFFICER I.D.: 0429

TESTING OFFICER:

GRIFFIN/B/C

OFFICER I.D.: 0429

PERMIT NUMBER: 240019

EXPIRATION DATE: 01/28/16

MISCELLANEOUS DATA:

N.F.L. TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	14:16
INTERNAL STANDARD	VERIFIED	14:17
ALCOHOL INTERFERENCE		

Operator Signature

BAC DataMaster Evidence Ticket

----- DIAGNOSTIC CHECK -----

[illegible][illegible]

Operator Signature

50